



Enzyme Research Laboratories

1801 Commerce Drive
South Bend, IN U.S.A. 46628
(574) 288-2268 * Fax (574) 288-2272
Website: www.enzymeresearch.com
Email: info@enzymeresearch.com

Matched-Pair Antibody Set for ELISA of Canine Factor VIII antigen (cFVIII)

Sufficient reagent for 4 x 96 well plates

Product #: CFVIII-EIA
Lot #: SAMPLE
Expiry Date: SAMPLE

Store at 2-8°C

For Research Use Only Not for use in
diagnostic procedures.

Description of Factor VIII (FVIII)

Factor VIII is a large glycoprotein (320 kDa) synthesized in the liver. The majority of Factor VIII is cleaved during expression, resulting in a mixture of partially cleaved forms ranging in size from 200-280 kDa. The FVIII is stabilized in circulation through non-covalent association with von Willebrand Factor. The concentration of FVIII in normal human plasma is typically 200 ng/mL. In canine plasma, FVIII activity is 5 to 7-fold higher relative to human plasma. FVIII is a pro-cofactor that is activated through limited proteolysis by thrombin. In this process FVIIIa dissociates from vWF to combine with activated Factor IX, calcium and a phospholipid surface where it is an essential cofactor in the assembly of the Factor X activator complex. Hemophilia A is a congenital bleeding disorder resulting from an X-chromosome-linked deficiency of FVIII, occurring with a frequency of 1 in 4000 males. The defect can be caused by any one of hundreds of reported mutations but are most commonly due to inversions within intron 22 of the FVIII gene. Hemophilia A has also been reported in a variety of species including dog and mouse, with a clinical phenotype very similar to human. The genetic defect in one case of canine Hemophilia-A has been shown to also be due to a gene inversion similar to the human defect, possibly indicating a common instability of the FVIII gene in humans and dogs¹⁻⁴.

Principle of Sandwich-style ELISA

Purified antibody to canine FVIII is coated onto the wells of a microtitre plate. The plate is washed and plasma or other fluids containing cFVIII are applied. The coated antibody will capture the cFVIII in the sample. After washing the plate to remove unbound material, a peroxidase conjugated second antibody to cFVIII is added to the plate to bind to the captured cFVIII. After washing the plate to remove unbound conjugated antibody, the peroxidase activity is expressed by incubation with o-phenylenediamine (OPD). After a fixed development time the reaction is quenched with the addition of H₂SO₄ and the colour produced is quantified using a microplate reader. The colour generated is proportional to the concentration of cFVIII present in the sample.

Supplied Materials:

- 1. Capture Antibody (CFVIII-EIA-C):** One yellow-capped vial containing 0.4 ml of polyclonal purified anti-canine FVIII antibody for coating plates.
- 2. Detecting Antibody (CFVIII-EIA-D):** Four neutral-capped tubes each containing 10 ml of pre-diluted peroxidase conjugated polyclonal anti-canine FVIII antibody for detection of captured cFVIII.
- 3. Sample Diluent (CFVIII-EIA-SD):** 100 ml bottle containing a green-coloured diluent optimised for dilution of samples.

Store reagents at 2-8°C

Materials Required but not Provided:

- 1. Coating Buffer:** 50 mM Carbonate 1.59g of Na₂CO₃ and 2.93g of NaHCO₃ up to 1 litre. Adjust pH to 9.6. Store at 2-8°C up to 1 month.
- 2. PBS:** (base for wash buffer) 8.0g NaCl, 1.15g Na₂HPO₄, 0.2g KH₂PO₄ and 0.2g KCl, up to 1 litre. Adjust pH to 7.4, if necessary. Store at 2-8°C up to 1 month, discard if there is evidence of microbial growth.
- 3. Wash Buffer:** PBS-Tween (0.1%,v/v) To 1 litre of PBS add 1.0 ml of Tween-20. Check that the pH is 7.4. Store at 2-8°C up to 1 week.
- 4. Substrate Buffer:** Citrate-Phosphate buffer pH 5.0 2.6g Citric acid and 6.9g Na₂HPO₄ up to a final volume of 500 ml with purified H₂O. Store at 2-8°C up to 1 month.
- 5. OPD Substrate:** (o-Phenylenediamine.2HCl) TOXIC! (5mg tablets: Sigma # P-6912). Make up immediately before use. Dissolve 5mg OPD in 12 ml substrate buffer then add 12 µl 30% H₂O₂. Do not store.
- 6. Stopping Solution:** 2.5 M H₂SO₄ Caution: VERY CORROSIVE! GENERATES HEAT ON DILUTION! Where stock sulphuric acid is 18 Molar, add 13.9 ml to 86 ml H₂O. Store at room temperature.

7. Other:

Microplates, 96-well Immulon 4-HBX (<http://www.labsystems.fi>)
Microplate washer (optional)
Microplate reader.

Assay Procedure:

1. Coating of plates: Dilute the capture antibody 1/100 in coating buffer (preferably in a polypropylene tube) and immediately add 100 µl to every well in the plate. Incubate 2 hrs @ 22°C or overnight @ 2-8°C.

2. Blocking: Blocking is not required under the conditions described. Washing the plate with PBS-Tween is sufficient to block nonspecific interactions. Wash plate X 3 with wash buffer.

3. Samples: To prepare a reference curve normal canine plasma is diluted 1/5 (100%) then serial 1/2's down to 1/160 (3.13%). Sample plasmas are diluted 1/10, 1/20 & 1/40. All dilutions are made in the provided green sample diluent. Apply 100 µl/well and incubate plate @ 22°C for 120 minutes. Wash plate X 3 with wash buffer.

4. Detecting Antibody: Apply the pre-diluted detecting antibody, 100 µl to each well. Incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

5. OPD Substrate: Apply 100 µl of freshly prepared OPD substrate to every well. Allow colour to develop for 5-10 minutes then stop colour reaction with the addition of 50 µl/well of 2.5 M H₂SO₄. The plate can be read at wavelength of 490 nm.

Calculation of Results:

The construction of a proper reference curve is of no less importance than any other aspect of the assay. A reference curve should be constructed by plotting the known concentration of standards versus absorbance. This can be done manually using graph paper, or by using curve-fitting computer software. In our experience, the dose response curves of most immunoassays tend to be sigmoid in shape. Although linear regions can be identified within the curve, the best overall fit is often obtained using an algorithm that provides a weighted theoretical model of fit throughout the entire curve, such as a 4-parameter or 5-parameter logistic curve fit^{5,6}. In general, the simplest model that defines the concentration-response relationship should be used⁷. The "back-fit" test is a simple and reliable method to determine if a curve-fitting method is appropriate. In this test, the apparent concentrations for the absorbance values of each standard point are read from the reference curve. The derived values are compared to the assigned values. An appropriate curve fitting method will produce derived values that closely match assigned values throughout the range of the curve, within user-defined

limits⁶. The coefficient of determination (R^2) is a valuable indicator of the overall fit, but should not be used by itself in the selection of a curve fitting method, as a poor fit in a particular region of the curve may not be evident from this value alone^{6,7}.

In the quality control of this product we have determined that under the conditions described above, a reference curve that is constructed using serial dilutions of normal canine plasma, will produce a correlation coefficient (R^2) of at least 0.980 using a log-log fit, and an R^2 of at least 0.990 using a 4-parameter logistic curve fit algorithm. However, the performance characteristics of in-house assays developed using this product in other laboratories may vary slightly from ours. Different curve fitting methods may be employed but we recommend that the back-fit test be applied as evidence that the fitting method is appropriate.

Technical Notes:

-This paired antibody product is intended to facilitate the end user in establishing an in-house immunoassay for research purposes only. Assay validation is the responsibility of the end user and should be done according to user-defined protocols⁷.

-Reference calibrators should be of the same matrix and anticoagulant as the samples to be tested (example serum or plasma, citrate or EDTA).

-The optimal colour development time should be determined empirically as the time required to obtain an absorbance of at least 1.000 at 490 nm for the 100% reference point, not to exceed 20 minutes.

-Do not use samples diluted less than 1/5, as falsely high readings may result.

-Rheumatoid factor or other heterophilic antibodies in samples may interfere in ELISA by binding to the capture and/or detecting antibodies.

-The wells should not be allowed to become dry. Keep plate covered or in a humid chamber during incubations.

-The capture antibody is supplied in a microcentrifuge tube and can be centrifuged to gather residual reagent from the cap and walls of the tube.

References:

1. Lollar P, Fay PJ, Fass DN; Factor VIII and Factor VIIIa. Methods in Enzymology, 222, pg 122, 1993.
2. Hoyer, LW, Wyshock EG, Colman RW, in Hemostasis and Thrombosis, 3rd Edition, eds. RW Colman, J Hirsh, VJ Marder and EW Salzman, pp. 109-133, J.B. Lippincott Co., Philadelphia, 1994.
3. Lozier JN, Dutra A. et al: The Chapel Hill hemophilia dog colony exhibits a factor VIII gene inversion. PNAS. 99, pg 12991-96, 2002.
4. Gallo-Penn AM, Shirley PS, Andrews JL, Tinlin S, Webster S, et al: Systemic delivery of an adenoviral vector encoding canine factor VIII results in short term phenotypic correction, inhibitor development and liver toxicity in hemophilia A dogs. Blood. 97, pg 107-113, 2001.
5. Nix,B, Wild D, in Immunoassays, A Practical Approach, editor J.P. Gosling, pp. 239-261, Oxford University Press, 2000.
6. NCCLS. Evaluation of the Linearity of Quantitative Analytical Methods; Proposed Guideline – Second Edition. NCCLS Document EP6-P2 (ISBN 1-56238-446-5, NCCLS, Wayne, Pennsylvania, 2001.
7. FDA Guidance for Industry. Bioanalytical Method Validation; May 2001, available online: www.fda.gov/cder/guidance/index.htm .

Related Products:

Cat #: SAF8C-IG	Sheep anti-human Factor VIII, IgG from antiserum
Cat #: SAF8C-AP	Sheep anti-human Factor VIII, affinity purified IgG
Cat #: SAF8C-HRP	Sheep anti-human Factor VIII, whole IgG-peroxidase
Cat #: SAF8C-APFTC	Sheep anti-human Factor VIII, FITC-conjugated APIgG
Cat #: SAF8C-APBIO	Sheep anti-human Factor VIII, biotinylated APIgG
Cat #: F8C-EIA	Paired antibodies for ELISA of human Factor VIII
Cat #: FVIII-DP	Human plasma deficient in FVIII, immune depleted
Cat #: FVIII-AG	Visualize Factor VIII Antigen Kit, for human IVD use